



RT-7600 for Vet

Auto Hematology Analyzer

شرکت مبنان پارس

فروش تجهیزات و لوازم آزمایشگاهی در سراسر کشور

مشاوره تاسیس آزمایشگاه، آموزش و راه اندازی بخش مولکولی و ژنتیک

آدرس: تهران، بزرگراه اشرفی اصفهانی جنب درمانگاه پیراه، پلاک ۵۴ واحد ۴

دفتر فروش: ۹-۱۵۸-۰۰۴۴-۰۲۱

مدیریت (نادر عدلی): ۰۳۰۷۰۳۴۱۲۱۳۴

www.mabnamed.com



Features

- 7 pre-determined animal settings plus 4 user-defined animal settings
- 3-part differentiation of WBC, 20 parameters+3 histograms
- High throughput, 60 samples per hour
- 10.4" TFT colorful LCD with touch screen
- External AC adapter avoids electronic nosies, enhancing accuracy
- Up to 50000 sample results (including histograms) can be stored
- User-friendly software, Multi-language available on request
- Original factory's blood control and calibrator
- Low sample & reagent consumption

Hiperion

Manufacture standard , Manufacture reliance

SCE

RT-7600 for Vet

Auto Hematology Analyzer



Animal Selection



Sample Test

Sample Profile

History Record List

Manual Calibration

Normal Value Setting



Technical Specifications

Animal Type:	Cat, Dog, Rabbit, Pig, Goat, Horse, Monkey plus 4 user-defined animal settings.
Parameters,	WBC, LYM#, MID#, GRA#, LYM%, MID%, GRA%, RBC, HGB, MCHC, MCH, MCV, RDW-CV, RDW-SD, HCT, PLT, MPV, PDW, PCT, P-LCR and Histograms for WBC, RBC and PLT
Principle,	Electrical resistance for counting WBC, RBC and PLT
Sample Volume,	Prediluted: 20µL; Whole Blood: 9.8µL
Throughput,	60 samples per hour
Display,	10.4" TFT color LCD with touch screen
Alarms,	Error messages
Input/Output,	RS-232, USB, LAN, keyboard and mouse Interface
Printout,	Thermal printer, 55mm width paper, various printout formats, external printer optional
Operating Environment:	Temperature: 15 ~35 °C; Humidity: ≤80%
Power Requirement,	AC 110V-220V, S0/60Hz
Dimension L*W*H (mm):	436*363*367
Weight,	18Kg

Hiperion

Hiperion

Medizintechnik GmbH & Co. KG
Carl-Zeiss-Str. 18 D-6332 Roedermark, Germany

Hiperion reserves the right to change specifications without prior notice. PIN.ENG-RT7600forVet-20170421